

Liquid Handling System Log

Version 2

Toxicology Unit

Effective Date: 08/29/2016

Liquid Handler Calibration Check – performed gravimetrically using deionized water

| | | | |
|---|--|---|--------------|
| Liquid Handler: | Application: <input type="checkbox"/> Volatiles <input type="checkbox"/> Immunoassay | Initials: | Date: |
| Comments/Maintenance: | | | |
| Check 1: Dispensed Total: _____g Sampled Total: _____g All Quality Criteria Met: <input type="checkbox"/> Yes <input type="checkbox"/> No | | Check 2: Dispensed Total: _____g Sampled Total: _____g All Quality Criteria Met: <input type="checkbox"/> Yes <input type="checkbox"/> No | |

Acceptable Ranges: Values must be within +/- 3% of target.

| | Volatiles | Immunoassay |
|---------------------|-----------------|-----------------|
| Dispensed Total (g) | 1.7460 - 1.8540 | 2.4250 - 2.5750 |
| Sampled Total (g) | 0.1940 - 0.2060 | 0.2425 - 0.2575 |