



### Certificate of Quality Control

### “A U k Y” Performance Check

Specifications:

The T æ, ^|| will be checked annually and after any repair or service with a known and a blank sample. Upon QPCR, the known must contain at least 0.0F5ng/µl of DNA and the blank must not have any DNA. If DNA is detected in the QPCR results for the blank, the test must be repeated by a DNA Team Leader or the Chief Criminalist.

Reason for the performance check: Annual

After repair, service, or preventative maintenance

Known Standard \_\_\_\_\_

Kit Used for Performance Check:

Lot # \_\_\_\_\_ Expiration Date \_\_\_\_\_ QC Date \_\_\_\_\_

Date Performance Check Completed \_\_\_\_\_ Analyst \_\_\_\_\_

Required records:

- Certificate of Quality Control
- QPCR results & setup worksheet
- DNA extraction worksheet

Results of Testing: Pass Fail (Choose One)

Comments:

Approval of COQC Record  
by Chief Criminalist, DNA  
Team Leader, or designee: \_\_\_\_\_

Date\* \_\_\_\_\_

\*If passed, this is the date released