

 <p align="center">Nonconformity Record Pitt County Sheriff's Office Forensics Services Unit <i>Issued by the Quality Manager</i></p>	Effective Date: 2018/04/01	Ver: 2
	Nonconformity Record	

NCR	19-010
<input type="checkbox"/>	Check if Customer Complaint

**Pitt County Forensic Services
NONCONFORMITY RECORD**

SECTION I

Individual and/or Laboratory	BAC Section
Type of forensic analysis	Blood Alcohol
Laboratory and/or Agency Case Number (if known/applicable)	
Date of Incident, Analysis, or Report (if known/applicable)	11/20/2019

SECTION II - Please write a *brief* statement of event(s).

BAC Technical procedures were not complete in this section upon audit. No Analyst or Tech Leader to complete

SECTION III - The person identifying the NCR:

Name	Chuck McClelland	Date	11/20/2019
------	------------------	------	------------

SECTION IV - Basic fact-finding. You may use additional paper, if necessary.

Criteria 7.6.1 states laboratories shall identify the contributions to measurement uncertainty. BAC Section was in the process of updating Technical Procedure to meet ISO 2017 Measurement Uncertainty standards when Chemist/ Tech Leader ended employment prior to completion.

Quality Manager/Technical Leader	Wallace Moore	Date	11/21/2019
Further Action Requested?	Yes	X	No

Explanation: All BAC operations were shut down after analyst left employment. No completed case work was subject to this NCR 19-010.